



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)

NABL / M-1336

15.04.2022

Wahid Ali

CHEMICAL PATHOLOGY LAB, POST GRADUATE DEPARTMENT OF PATHOLOGY, KING GEORGE'S MEDICAL UNIVERSITY
KGMU, SHAHIMEENA ROAD CHOWK
LUCKNOW, UTTAR PRADESH-226003
Mobile: 9198852343
E-mail: pgdp.kgm@gmail.com

Subject: **Renewal of accreditation as per ISO 15189:2012**

Dear Sir,

In reference to the reassessment of your laboratory, NABL is pleased to inform the renewal of accreditation in accordance with ISO 15189:2012 in the field of Medical testing for the discipline of Clinical Biochemistry as per the scope recommended by the assessment team. However, following is also advised which the laboratory has to address and same to be verified in the next assessment:

- Disclaimer in the test reports of LDL-Cholesterol as "Estimation of LDL-Cholesterol by direct method is recommended when Triglyceride > 400 mg/dl".

All the personnel proposed by the laboratory to review the results and authorize the release of reports of Clinical Biochemistry are accepted.

With respect to the persons to review the results and authorize the release of reports, it is the responsibility of laboratory to abide by the National/ Regional/ State/ Local regulatory requirements/ Acts/ Rules/ Legal orders/ Court Decisions/ Orders issued by Government/ Statutory Bodies as applicable and effective from time to time.

The accreditation certificate no. is MC-2563 issue date 30.01.2022 valid till 29.01.2024, Amendment date: 13.04.2022. The certificate is in process of making and it would be sent to you soon.

NABL is now allowing its accredited CABs (testing, calibration and medical laboratories) to use 'NABL Accredited CAB Combined ILAC MRA Mar' on their test /calibration reports through a valid Agreement. For more details, please refer NABL document NABL 133 available on NABL website www.nabl-india.org under publications section.

Yours Sincerely,

Jayeeta Majumdar

jayeeta@nabl.qcin.org